
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006702
12:25 06/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:29
PRESSURIZING 1:32
STERILIZING 2:29
136.4 C 220kPa 2:29
136.0 C 222kPa 2:59
136.4 C 224kPa 3:29
136.4 C 224kPa 3:59
136.3 C 223kPa 4:29
136.3 C 217kPa 4:59
136.0 C 221kPa 5:29
136.4 C 222kPa 5:59
136.4 C 222kPa 5:59
VENTING 5:59
AIR DRYING 6:48
CYCLE COMPLETE 6:58

Digital Signature #
106353569C470482
