

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1857

Name: RITTER2  
Sterilizer ID: V2313228  
12/19/2022 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.9 F 0.0  
2:00 200.4 F 4.7  
4:00 208.5 F 10.6  
6:00 223.5 F 6.8  
8:00 237.4 F 13.2  
10:00 240.7 F 11.9  
12:00 241.6 F 11.5  
14:00 255.8 F 19.0  
16:00 267.9 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.5 F 29.6  
2:00 271.7 F 29.3  
3:00 272.6 F 29.6  
4:00 271.6 F 29.0  
  
Min 271.6 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 225.1 F 0.0  
10:00 241.2 F 0.0  
15:00 228.9 F 0.0  
20:00 224.6 F 0.0  
25:00 222.7 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 2:55  
HEATING: 17:13  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:18  
CYCLE COMPLETE

12/19/2022 03:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_