

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2785

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2023 01:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.2 F 0.0  
2:00 169.5 F 2.4  
4:00 173.1 F 5.8  
6:00 202.6 F 10.8  
8:00 219.8 F 4.8  
10:00 229.4 F 9.8  
12:00 234.8 F 8.7  
14:00 242.1 F 12.2  
16:00 240.1 F 10.7  
18:00 253.5 F 17.3  
20:00 265.1 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.5 F 29.3  
2:00 271.5 F 28.4  
3:00 272.6 F 29.2  
4:00 271.7 F 28.5  
  
Min 271.4 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 225.6 F 0.0  
10:00 232.9 F 0.0  
15:00 227.7 F 0.0  
20:00 225.8 F 0.0  
25:00 228.6 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 21:39  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

01/17/2023 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_