

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4391

Name: RITTER 1  
Sterilizer ID: V2178198  
03/21/2024 11:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.2 F 0.0  
2:00 114.7 F 1.0  
4:00 126.2 F 3.4  
6:00 155.7 F 6.1  
8:00 185.0 F 9.8  
10:00 206.7 F 9.8  
12:00 211.6 F 5.5  
14:00 222.0 F 9.2  
16:00 232.0 F 13.3  
18:00 233.7 F 9.2  
20:00 241.4 F 11.7  
22:00 242.7 F 11.9  
24:00 253.7 F 17.5  
26:00 263.7 F 23.4  
28:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.1 F 29.1  
2:00 272.1 F 28.8  
3:00 271.4 F 28.7  
4:00 272.5 F 29.2

Min 271.2 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 222.3 F 0.0  
10:00 231.4 F 0.0  
15:00 233.3 F 0.0  
20:00 222.8 F 0.0  
25:00 222.2 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 28:10  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:06:56  
CYCLE COMPLETE

03/21/2024 12:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_