
Midmark M11 - v1.0.5
Total Cycles: 3619

Name: RITTER2
Sterilizer ID: V2313228
05/15/2024 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 151.7 F 1.5
4:00 152.2 F 4.1
6:00 167.0 F 7.6
8:00 188.7 F 11.5
10:00 210.4 F 4.3
12:00 220.3 F 8.4
14:00 231.7 F 12.7
16:00 233.3 F 9.4
18:00 239.3 F 10.6
20:00 243.9 F 13.1
22:00 254.7 F 18.8
24:00 264.8 F 24.8
26:00 271.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.2 F 30.3
2:00 272.5 F 30.0
3:00 271.7 F 29.6
4:00 271.2 F 29.4

Min 271.1 F 29.3
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 225.1 F 0.0
10:00 223.6 F 0.0
15:00 213.0 F 0.0
20:00 223.6 F 0.0
25:00 230.8 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:11
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:00
CYCLE COMPLETE

05/15/2024 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____