
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000061
7:39 30/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:53
STERILIZING 3:55
136.5 C 220kPa 3:55
135.8 C 220kPa 4:25
136.4 C 223kPa 4:55
136.4 C 221kPa 5:25
136.2 C 223kPa 5:55
136.1 C 219kPa 6:25
136.3 C 222kPa 6:55
135.9 C 221kPa 7:25
135.9 C 221kPa 7:25
VENTING 7:25
AIR DRYING 8:14
CYCLE COMPLETE 8:20

Digital Signature #
56A9C09B0412DB98
