
Midmark M11 - v1.0.5
Total Cycles: 4319

Name: RITTER2
Sterilizer ID: V2313228
01/30/2025 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.5 F 0.0
2:00 134.1 F 1.5
4:00 151.5 F 4.9
6:00 176.3 F 9.1
8:00 201.1 F 11.0
10:00 220.4 F 6.6
12:00 235.0 F 11.1
14:00 229.5 F 7.3
16:00 243.5 F 12.9
18:00 240.8 F 11.3
20:00 253.3 F 17.5
22:00 264.3 F 24.1
24:00 271.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 273.1 F 29.7
2:00 271.7 F 29.7
3:00 272.3 F 29.6
4:00 271.5 F 29.8

Min 271.1 F 29.2
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 233.2 F 0.0
10:00 214.2 F 0.0
15:00 206.5 F 0.0
20:00 207.5 F 0.0
25:00 194.4 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

01/30/2025 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____