
Midmark M11 - v1.0.0
Total Cycles: 867

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2021 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 175.6 F 3.2
4:00 193.6 F 7.9
6:00 221.6 F 9.5
8:00 234.3 F 9.6
10:00 233.5 F 8.1
12:00 244.9 F 12.6
14:00 245.7 F 12.9
16:00 259.1 F 20.1
18:00 271.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.6
1:00 272.6 F 28.9
2:00 271.8 F 28.7
3:00 272.5 F 29.0
4:00 272.4 F 28.7

Min 271.8 F 28.5
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 233.9 F 0.0
10:00 225.1 F 0.0
15:00 231.7 F 0.0
20:00 229.9 F 0.0
25:00 231.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

06/07/2021 01:34 PM
ACCEPTED: X
REJECTED:
Operator: U01