
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001550
10:41 11/03/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:53
PRESSURIZING 6:39
STERILIZING 8:00
136.0 C 220kPa 8:00
136.4 C 219kPa 8:30
136.1 C 219kPa 9:00
135.3 C 216kPa 9:30
135.6 C 217kPa 10:00
136.5 C 222kPa 10:30
136.5 C 222kPa 11:00
136.1 C 220kPa 11:30
VENTING 11:30
AIR DRYING 12:34
CYCLE COMPLETE 42:34

Digital Signature #
1EFB45F550B6B85A
