
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002991
15:38 19/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:54
STERILIZING 2:52
136.4 C 224kPa 2:52
136.3 C 224kPa 3:22
136.3 C 223kPa 3:52
136.2 C 221kPa 4:22
136.4 C 224kPa 4:52
136.4 C 222kPa 5:22
136.4 C 224kPa 5:52
136.4 C 224kPa 6:22
136.4 C 224kPa 6:22
VENTING 6:22
AIR DRYING 7:17
CYCLE COMPLETE 7:34

Digital Signature #
F1B5EF6507452EFD
