

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4047

Name: RITTER2  
Sterilizer ID: V2313228  
10/23/2024 08:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 89.3 F 0.0  
4:00 130.4 F 0.0  
6:00 127.7 F 0.1  
8:00 143.4 F 0.5  
10:00 175.1 F 2.3  
12:00 201.4 F 6.1  
14:00 210.5 F 10.2  
16:00 227.1 F 7.9  
18:00 230.8 F 8.7  
20:00 242.6 F 13.4  
22:00 239.3 F 10.8  
24:00 236.3 F 9.1  
26:00 250.4 F 15.9  
28:00 261.9 F 22.6  
30:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.1 F 28.9  
2:00 272.4 F 29.3  
3:00 272.7 F 29.6  
4:00 271.9 F 29.3

Min 271.7 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
5:00 223.8 F 0.0  
10:00 223.3 F 0.0  
15:00 209.1 F 0.0  
20:00 193.2 F 0.0  
25:00 181.5 F 0.0  
30:00 178.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:20  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:23  
CYCLE COMPLETE

10/23/2024 09:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_