
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006712
9:40 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:27
136.0 C 220kPa 3:27
136.3 C 222kPa 3:57
136.1 C 221kPa 4:27
136.4 C 224kPa 4:57
136.3 C 223kPa 5:27
136.0 C 220kPa 5:57
136.3 C 223kPa 6:27
136.1 C 219kPa 6:57
136.1 C 219kPa 6:57
VENTING 6:57
AIR DRYING 7:43
CYCLE COMPLETE 11:53

Digital Signature #
21DCA6DA5F0BB821
