

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1349

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2022 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
2:00 188.7 F 0.4  
4:00 210.2 F 5.0  
6:00 225.9 F 10.4  
8:00 230.8 F 6.5  
10:00 238.0 F 12.2  
12:00 238.0 F 10.4  
14:00 238.1 F 9.5  
16:00 251.5 F 16.5  
18:00 263.3 F 23.4  
20:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.6 F 29.5  
2:00 271.7 F 29.1  
3:00 272.6 F 29.5  
4:00 271.7 F 28.9

Min 271.6 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 213.1 F 0.0  
10:00 222.5 F 0.0  
15:00 225.2 F 0.0  
20:00 229.9 F 0.0  
Dry Cycle Aborted

FILLING: 2:44  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 23:54  
TOTAL CYCLE: 00:53:57

07/26/2022 02:15 PM

ACCEPTED:

REJECTED:

Reason: -----  
Operator: -----