

-----  
Midmark M11 - v1.0.5  
Total Cycles: 246

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2021 11:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.6 F 0.0  
2:00 154.4 F 2.0  
4:00 167.6 F 5.8  
6:00 190.4 F 11.0  
8:00 215.8 F 5.8  
10:00 227.4 F 10.9  
12:00 229.9 F 7.8  
14:00 242.8 F 13.5  
16:00 242.6 F 12.2  
18:00 254.6 F 18.5  
20:00 265.7 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.6 F 29.8  
2:00 271.5 F 29.2  
3:00 272.6 F 29.9  
4:00 271.9 F 29.3  
  
Min 271.4 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 219.3 F 0.0  
10:00 220.3 F 0.0  
15:00 222.7 F 0.0  
20:00 221.8 F 0.0  
25:00 225.6 F 0.0  
30:00 178.6 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:15  
CYCLE COMPLETE

07/26/2021 12:44 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01