
Midmark M11 - v1.0.0
Total Cycles: 1775

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2022 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 190.7 F 3.4
4:00 206.4 F 7.9
6:00 226.6 F 10.5
8:00 232.0 F 9.3
10:00 238.6 F 9.2
12:00 243.5 F 12.9
14:00 242.4 F 11.8
16:00 255.7 F 18.6
18:00 267.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.1
2:00 271.5 F 28.5
3:00 272.4 F 29.2
4:00 271.7 F 28.5

Min 271.4 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 216.4 F 0.0
10:00 235.7 F 0.0
15:00 234.0 F 0.0
20:00 225.3 F 0.0
25:00 222.5 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

03/31/2022 10:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____