

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2653

Name: RITTER2  
Sterilizer ID: V2313228  
07/17/2023 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
2:00 187.5 F 2.1  
4:00 194.4 F 6.4  
6:00 210.6 F 10.6  
8:00 220.7 F 6.5  
10:00 231.1 F 9.4  
12:00 237.5 F 10.9  
14:00 241.7 F 12.2  
16:00 241.1 F 11.2  
18:00 253.0 F 17.3  
20:00 263.5 F 23.5  
22:00 272.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.3 F 29.5  
2:00 271.2 F 28.9  
3:00 272.3 F 29.6  
4:00 271.8 F 29.0

Min 271.1 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 214.9 F 0.0  
10:00 213.5 F 0.0  
15:00 207.8 F 0.0  
20:00 205.6 F 0.0  
25:00 209.1 F 0.0  
30:00 212.8 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 22:15  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:52  
CYCLE COMPLETE

07/17/2023 12:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_