

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000101  
11:29 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:32  
PRESSURIZING 4:33  
STERILIZING 5:57  
136.0 C 220kPa 5:57  
136.0 C 224kPa 6:27  
136.4 C 227kPa 6:57  
136.3 C 226kPa 7:27  
136.2 C 225kPa 7:57  
136.4 C 226kPa 8:27  
136.4 C 225kPa 8:57  
136.4 C 227kPa 9:27  
136.4 C 227kPa 9:27  
VENTING 9:27  
AIR DRYING 10:27  
CYCLE COMPLETE 25:27  
-----

Digital Signature #  
850F885BCCA06889  
-----