

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2616

Name: RITTER 1  
Sterilizer ID: V2178198  
11/28/2022 04:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
2:00 208.7 F 4.7  
4:00 218.2 F 10.3  
6:00 227.9 F 5.8  
8:00 238.0 F 11.8  
10:00 237.5 F 9.6  
12:00 238.6 F 8.6  
14:00 250.5 F 15.6  
16:00 263.2 F 22.8  
18:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 271.9 F 28.7  
2:00 271.5 F 28.7  
3:00 271.9 F 28.7  
4:00 271.3 F 28.5  
  
Min 271.2 F 28.2  
Max 272.3 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 217.0 F 0.0  
5:00 221.4 F 0.0  
10:00 246.1 F 0.0  
15:00 237.9 F 0.0  
20:00 226.7 F 0.0  
25:00 228.8 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 3:10  
HEATING: 18:03  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:26  
CYCLE COMPLETE

11/28/2022 05:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_