
Midmark M11 - v1.0.0
Total Cycles: 4497

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2024 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.8 F 0.0
2:00 120.8 F 1.2
4:00 139.8 F 3.6
6:00 164.5 F 6.6
8:00 178.6 F 10.2
10:00 221.5 F 6.6
12:00 227.2 F 7.0
14:00 234.2 F 11.1
16:00 230.0 F 6.2
18:00 239.4 F 11.5
20:00 236.8 F 8.7
22:00 249.2 F 14.5
24:00 260.1 F 20.6
26:00 270.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.6
2:00 272.7 F 29.0
3:00 271.7 F 29.1
4:00 271.6 F 28.5

Min 271.2 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.7 F 0.0
5:00 235.9 F 0.0
10:00 189.7 F 0.0
15:00 187.8 F 0.0
20:00 187.5 F 0.0
25:00 187.8 F 0.0
30:00 186.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 27:39
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

04/26/2024 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____