

-----  
Midmark M11 - v1.0.5  
Total Cycles: 518

Name: RITTER2  
Sterilizer ID: V2313228  
11/03/2021 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
2:00 172.9 F 3.1  
4:00 173.5 F 7.5  
6:00 195.5 F 13.2  
8:00 214.7 F 8.4  
10:00 230.7 F 10.5  
12:00 238.2 F 12.0  
14:00 239.5 F 10.9  
16:00 253.3 F 17.9  
18:00 265.0 F 25.1  
20:00 270.8 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.9  
1:00 272.8 F 30.1  
2:00 272.4 F 30.6  
3:00 272.2 F 29.9  
4:00 272.9 F 30.5

Min 271.8 F 29.8  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.9 F 0.0  
5:00 228.5 F 0.0  
10:00 228.1 F 0.0  
15:00 226.9 F 0.0  
20:00 223.9 F 0.0  
25:00 225.9 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

11/03/2021 02:46 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01