

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2194

Name: RITTER2  
Sterilizer ID: V2313228  
03/29/2023 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.5 F 0.0  
2:00 135.2 F 2.3  
4:00 158.2 F 5.9  
6:00 187.0 F 11.0  
8:00 208.8 F 4.7  
10:00 224.5 F 9.9  
12:00 229.7 F 8.2  
14:00 240.7 F 12.6  
16:00 240.4 F 11.3  
18:00 254.0 F 18.2  
20:00 265.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.8 F 29.7  
2:00 272.0 F 29.6  
3:00 272.7 F 29.6  
4:00 271.9 F 29.5

Min 271.6 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.5 F 0.0  
5:00 206.8 F 0.0  
10:00 222.9 F 0.0  
15:00 230.1 F 0.0  
20:00 227.6 F 0.0  
25:00 228.5 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:37  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

03/29/2023 02:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_