
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004432
10:13 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 2:58
STERILIZING 4:03
136.4 C 221kPa 4:03
136.4 C 224kPa 4:33
136.4 C 223kPa 5:03
136.4 C 222kPa 5:33
136.2 C 218kPa 6:03
136.4 C 223kPa 6:33
136.4 C 222kPa 7:03
136.4 C 221kPa 7:33
136.4 C 221kPa 7:33
VENTING 7:33
AIR DRYING 8:25
CYCLE COMPLETE 8:31

Digital Signature #
453A644DE5D0D710
