

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2244

Name: RITTER 1  
Sterilizer ID: V2178198  
08/25/2022 07:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 76.7 F 0.0  
4:00 85.1 F 0.0  
6:00 110.2 F 0.2  
8:00 142.5 F 0.5  
10:00 176.5 F 0.8  
12:00 199.9 F 2.0  
14:00 221.8 F 4.9  
16:00 228.8 F 7.8  
18:00 237.9 F 11.4  
20:00 231.7 F 6.4  
22:00 242.2 F 11.8  
24:00 237.7 F 9.3  
26:00 239.6 F 9.4  
28:00 248.1 F 14.0  
30:00 259.5 F 20.2  
32:00 270.1 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.3 F 28.3  
2:00 272.2 F 28.6  
3:00 271.8 F 28.0  
4:00 272.7 F 28.7  
  
Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 229.6 F 0.0  
10:00 200.3 F 0.0  
15:00 220.3 F 0.0  
20:00 230.7 F 0.0  
25:00 228.9 F 0.0  
30:00 226.2 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:52  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:10:53  
CYCLE COMPLETE

08/25/2022 08:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_