
Midmark M11 - v1.0.5
Total Cycles: 1787

Name: RITTER2
Sterilizer ID: V2313228
11/29/2022 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.3 F 0.0
2:00 169.1 F 0.2
4:00 191.8 F 3.0
6:00 211.6 F 7.8
8:00 228.9 F 13.2
10:00 234.8 F 10.1
12:00 235.2 F 8.5
14:00 241.2 F 10.6
16:00 248.8 F 14.8
18:00 261.5 F 22.2
20:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.5 F 29.6
2:00 271.9 F 28.8
3:00 272.4 F 29.6
4:00 272.2 F 29.1

Min 271.5 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 215.5 F 0.0
10:00 225.5 F 0.0
15:00 235.3 F 0.0
20:00 230.1 F 0.0
25:00 223.7 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

11/29/2022 05:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____