
Midmark M11 - v1.0.0
Total Cycles: 2591

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 87.3 F 0.0
4:00 143.9 F 0.1
6:00 164.7 F 1.7
8:00 178.1 F 4.5
10:00 201.8 F 8.4
12:00 216.5 F 12.6
14:00 219.1 F 4.9
16:00 229.2 F 8.8
18:00 238.8 F 12.9
20:00 235.9 F 9.3
22:00 237.9 F 9.6
24:00 245.9 F 12.8
26:00 257.3 F 19.0
28:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.9 F 29.0
2:00 272.5 F 28.8
3:00 271.8 F 28.5
4:00 271.2 F 28.7

Min 271.2 F 28.5
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 232.4 F 0.0
10:00 227.3 F 0.0
15:00 233.4 F 0.0
20:00 223.1 F 0.0
25:00 224.0 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

11/22/2022 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____