
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002265
8:46 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:32
STERILIZING 3:36
136.0 C 220kPa 3:36
136.2 C 223kPa 4:06
136.1 C 224kPa 4:36
136.4 C 224kPa 5:06
136.4 C 224kPa 5:36
136.3 C 223kPa 6:06
136.4 C 222kPa 6:36
136.1 C 221kPa 7:06
136.1 C 221kPa 7:06
VENTING 7:06
AIR DRYING 7:57
CYCLE COMPLETE 21:32

Digital Signature #
70DBCA1FE509208B
