

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1965

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2023 12:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.1 F 0.0  
2:00 132.4 F 0.1  
4:00 152.5 F 0.3  
6:00 196.6 F 0.7  
8:00 212.5 F 4.5  
10:00 223.6 F 8.9  
12:00 234.3 F 11.4  
14:00 238.0 F 10.9  
16:00 236.1 F 9.2  
18:00 238.6 F 9.6  
20:00 250.3 F 15.6  
22:00 262.6 F 22.7  
24:00 272.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.2  
2:00 271.5 F 28.6  
3:00 272.7 F 29.4  
4:00 271.9 F 28.7

Min 271.4 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 227.3 F 0.0  
10:00 232.0 F 0.0  
15:00 228.0 F 0.0  
20:00 231.6 F 0.0  
25:00 228.8 F 0.0  
30:00 229.7 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:12  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:22  
CYCLE COMPLETE

01/24/2023 01:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_