
Midmark M11 - v1.0.5
Total Cycles: 3522

Name: RITTER2
Sterilizer ID: V2313228
04/12/2024 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.0 F 0.0
2:00 143.6 F 1.7
4:00 160.4 F 4.6
6:00 184.1 F 8.3
8:00 201.0 F 12.2
10:00 214.8 F 5.6
12:00 225.5 F 9.9
14:00 231.9 F 9.7
16:00 238.3 F 11.6
18:00 237.3 F 10.2
20:00 249.8 F 16.1
22:00 260.3 F 22.3
24:00 270.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.5 F 30.1
2:00 271.3 F 29.6
3:00 272.5 F 30.2
4:00 271.6 F 29.6

Min 271.3 F 29.5
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 224.2 F 0.0
10:00 199.3 F 0.0
15:00 194.8 F 0.0
20:00 191.5 F 0.0
25:00 188.5 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:52
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

04/12/2024 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____