
Midmark M11 - v1.0.5
Total Cycles: 1222

Name: RITTER2
Sterilizer ID: V2313228
06/24/2022 02:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 181.0 F 3.9
4:00 191.4 F 8.5
6:00 215.1 F 11.8
8:00 225.0 F 9.2
10:00 234.8 F 9.6
12:00 241.0 F 12.6
14:00 241.1 F 11.6
16:00 254.0 F 18.2
18:00 265.3 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.8
2:00 271.6 F 29.1
3:00 272.5 F 30.0
4:00 272.1 F 29.5

Min 271.3 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 216.7 F 0.0
10:00 232.7 F 0.0
15:00 231.0 F 0.0
20:00 225.5 F 0.0
25:00 222.5 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

06/24/2022 03:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____