
Midmark M11 - v1.0.0
Total Cycles: 4526

Name: RITTER 1
Sterilizer ID: V2178198
05/06/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 123.1 F 1.2
4:00 130.3 F 3.4
6:00 158.1 F 6.2
8:00 178.4 F 9.5
10:00 195.7 F 11.4
12:00 212.2 F 6.4
14:00 225.0 F 10.6
16:00 229.6 F 7.5
18:00 237.8 F 11.5
20:00 235.8 F 9.0
22:00 247.8 F 14.5
24:00 257.9 F 20.1
26:00 267.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.0
1:00 272.2 F 29.0
2:00 272.3 F 29.0
3:00 272.3 F 29.0
4:00 272.1 F 29.1

Min 271.7 F 28.9
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 215.3 F 0.0
10:00 220.9 F 0.0
15:00 227.4 F 0.0
20:00 222.9 F 0.0
25:00 225.9 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 27:14
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:18
CYCLE COMPLETE

05/06/2024 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____