

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2882

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2023 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.3 F 0.0  
2:00 181.0 F 4.3  
4:00 191.6 F 8.9  
6:00 206.9 F 9.8  
8:00 222.0 F 7.7  
10:00 237.3 F 12.9  
12:00 237.2 F 10.1  
14:00 236.3 F 8.2  
16:00 249.3 F 15.1  
18:00 261.4 F 21.9  
20:00 272.3 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.5 F 29.4  
2:00 272.4 F 29.0  
3:00 271.6 F 28.5  
4:00 271.6 F 28.8  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 218.4 F 0.0  
10:00 232.7 F 0.0  
15:00 224.8 F 0.0  
20:00 223.7 F 0.0  
25:00 225.6 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 20:33  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

02/09/2023 02:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_