
Midmark M11 - v1.0.5
Total Cycles: 2791

Name: RITTER2
Sterilizer ID: V2313228
08/18/2023 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 156.4 F 0.0
4:00 175.7 F 0.0
6:00 202.2 F 0.0
8:00 209.8 F 0.5
10:00 223.9 F 4.8
12:00 234.2 F 9.0
14:00 244.1 F 12.7
16:00 242.8 F 12.2
18:00 241.7 F 11.5
20:00 241.3 F 11.3
22:00 254.6 F 18.0
24:00 265.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.3
2:00 271.6 F 28.6
3:00 272.8 F 29.5
4:00 272.0 F 28.9

Min 271.5 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 236.9 F 0.0
10:00 216.9 F 0.0
15:00 212.7 F 0.0
20:00 192.8 F 0.0
25:00 192.8 F 0.0
30:00 185.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:14
CYCLE COMPLETE

08/18/2023 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____