
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000666
9:01 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:48
STERILIZING 3:57
136.0 C 227kPa 3:57
136.3 C 231kPa 4:27
136.4 C 231kPa 4:57
136.3 C 231kPa 5:27
136.3 C 231kPa 5:57
136.3 C 231kPa 6:27
136.2 C 231kPa 6:57
136.4 C 232kPa 7:27
136.4 C 232kPa 7:27
VENTING 7:27
AIR DRYING 8:21
CYCLE COMPLETE 8:24

Digital Signature #
FCE929A076D4B743
