

-----  
Midmark M11 - v1.0.5  
Total Cycles: 494

Name: RITTER2  
Sterilizer ID: V2313228  
10/28/2021 06:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 74.0 F 0.0  
4:00 82.6 F 0.0  
6:00 111.0 F 0.2  
8:00 152.1 F 0.3  
10:00 183.4 F 2.1  
12:00 202.1 F 5.7  
14:00 213.8 F 9.5  
16:00 224.7 F 13.4  
18:00 230.3 F 8.3  
20:00 241.0 F 12.9  
22:00 239.4 F 10.6  
24:00 236.1 F 8.9  
26:00 250.4 F 15.7  
28:00 262.1 F 22.5  
30:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.5 F 29.1  
2:00 271.8 F 29.0  
3:00 272.1 F 29.0  
4:00 271.8 F 29.0  
  
Min 271.4 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 226.8 F 0.0  
10:00 230.3 F 0.0  
15:00 208.3 F 0.0  
20:00 215.5 F 0.0  
25:00 215.4 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 30:23  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:08:23  
CYCLE COMPLETE

10/28/2021 07:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_