
Midmark M11 - v1.0.5
Total Cycles: 3428

Name: RITTER2
Sterilizer ID: V2313228
03/01/2024 03:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.7 F 0.0
2:00 102.6 F 0.0
4:00 112.7 F 0.6
6:00 142.2 F 2.9
8:00 168.2 F 6.1
10:00 189.0 F 9.7
12:00 204.5 F 13.5
14:00 217.7 F 6.8
16:00 230.4 F 11.0
18:00 230.0 F 7.2
20:00 242.2 F 12.7
22:00 241.0 F 11.1
24:00 252.5 F 17.0
26:00 263.1 F 23.1
28:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.0 F 29.3
2:00 272.2 F 29.4
3:00 272.7 F 29.6
4:00 272.2 F 29.8

Min 271.6 F 29.3
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 225.3 F 0.0
10:00 205.4 F 0.0
15:00 198.5 F 0.0
20:00 194.7 F 0.0
25:00 192.1 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:06
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:36
CYCLE COMPLETE

03/01/2024 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____