
Midmark M11 - v1.0.0
Total Cycles: 783

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2021 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.3 F 0.0
2:00 123.1 F 0.5
4:00 132.5 F 2.4
6:00 160.4 F 5.7
8:00 193.2 F 10.3
10:00 227.7 F 5.6
12:00 226.3 F 10.6
14:00 231.2 F 6.5
16:00 245.4 F 12.9
18:00 241.9 F 11.0
20:00 255.1 F 17.7
22:00 266.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 271.9 F 28.7
2:00 272.6 F 29.1
3:00 272.5 F 28.9
4:00 272.4 F 28.8

Min 271.0 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 240.5 F 0.0
10:00 226.9 F 0.0
15:00 229.9 F 0.0
20:00 226.1 F 0.0
25:00 229.1 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

04/14/2021 11:50 AM
ACCEPTED: X
REJECTED:
Operator: U01