
Midmark M11 - v1.0.5
Total Cycles: 66

Name: RITTER2
Sterilizer ID: V2313228
06/01/2021 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.8 F 0.0
2:00 113.4 F 1.0
4:00 127.0 F 3.3
6:00 151.5 F 6.4
8:00 178.0 F 10.0
10:00 198.9 F 9.8
12:00 212.0 F 5.5
14:00 223.2 F 9.7
16:00 233.2 F 10.2
18:00 238.3 F 11.5
20:00 238.5 F 10.2
22:00 250.3 F 16.1
24:00 260.8 F 22.2
26:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.3 F 29.8
2:00 272.6 F 29.7
3:00 271.7 F 29.2
4:00 272.7 F 30.1

Min 271.7 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 218.8 F 0.0
10:00 226.6 F 0.0
15:00 220.3 F 0.0
20:00 216.9 F 0.0
25:00 210.2 F 0.0
30:00 208.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:05:26
CYCLE COMPLETE

06/01/2021 05:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____