
Midmark M11 - v1.0.5
Total Cycles: 2321

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 176.8 F 3.6
4:00 183.1 F 8.0
6:00 205.9 F 13.5
8:00 223.1 F 8.6
10:00 233.9 F 11.1
12:00 238.8 F 11.9
14:00 239.4 F 10.9
16:00 252.5 F 17.7
18:00 264.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.5 F 30.3
2:00 271.5 F 29.6
3:00 272.3 F 30.5
4:00 272.3 F 30.0

Min 271.3 F 29.5
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 210.3 F 0.0
10:00 230.4 F 0.0
15:00 225.7 F 0.0
20:00 224.2 F 0.0
25:00 222.9 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

05/01/2023 03:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____