
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001226
9:29 22/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:32
STERILIZING 5:09
136.4 C 225kPa 5:09
136.3 C 223kPa 5:39
136.2 C 223kPa 6:09
136.4 C 225kPa 6:39
136.4 C 224kPa 7:09
136.4 C 223kPa 7:39
136.2 C 222kPa 8:09
136.4 C 221kPa 8:39
136.4 C 221kPa 8:39
VENTING 8:39
AIR DRYING 9:34
CYCLE COMPLETE 24:10

Digital Signature #
98E5FFFAC9B2E952
