
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002155
9:40 17/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:42
136.4 C 224kPa 3:42
136.4 C 225kPa 4:12
136.3 C 224kPa 4:42
136.4 C 225kPa 5:12
136.3 C 223kPa 5:42
136.2 C 223kPa 6:12
136.3 C 223kPa 6:42
136.4 C 224kPa 7:12
136.4 C 224kPa 7:12
VENTING 7:12
AIR DRYING 8:01
CYCLE COMPLETE 8:05

Digital Signature #
FC4F775582CDBB8B
