
Midmark M11 - v1.0.5
Total Cycles: 1676

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 160.7 F 1.0
4:00 194.2 F 4.7
6:00 205.7 F 10.2
8:00 222.6 F 4.8
10:00 232.3 F 10.2
12:00 236.0 F 7.7
14:00 242.8 F 13.2
16:00 243.4 F 12.1
18:00 255.6 F 18.8
20:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.4 F 29.6
2:00 271.2 F 29.0
3:00 272.4 F 29.7
4:00 271.8 F 29.1

Min 271.0 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 207.5 F 0.0
10:00 227.8 F 0.0
15:00 232.0 F 0.0
20:00 226.9 F 0.0
25:00 227.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

11/03/2022 09:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____