
Midmark M11 - v1.0.0
Total Cycles: 3339

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.8 F 0.0
2:00 76.2 F 0.0
4:00 132.9 F 0.0
6:00 177.7 F 0.0
8:00 157.7 F 0.0
10:00 186.9 F 0.9
12:00 197.5 F 4.4
14:00 212.6 F 7.3
16:00 222.2 F 10.6
18:00 228.1 F 7.9
20:00 234.3 F 9.6
22:00 238.9 F 10.5
24:00 242.7 F 11.9
26:00 239.7 F 10.2
28:00 252.8 F 16.6
30:00 263.9 F 23.1
32:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.1
3:00 272.6 F 28.9
4:00 271.8 F 28.1

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 202.7 F 0.0
10:00 219.2 F 0.0
15:00 221.2 F 0.0
20:00 223.8 F 0.0
25:00 228.1 F 0.0
30:00 229.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

06/06/2023 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____