
Midmark M11 - v1.0.5
Total Cycles: 2209

Name: RITTER2
Sterilizer ID: V2313228
04/03/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 165.5 F 3.5
4:00 184.9 F 7.8
6:00 209.3 F 13.1
8:00 225.3 F 7.4
10:00 238.9 F 13.0
12:00 239.3 F 11.1
14:00 238.8 F 10.2
16:00 253.3 F 17.4
18:00 265.0 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.1 F 29.4
2:00 271.5 F 29.2
3:00 272.1 F 29.4
4:00 271.3 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 231.0 F 0.0
10:00 219.5 F 0.0
15:00 223.1 F 0.0
20:00 202.2 F 0.0
25:00 198.2 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

04/03/2023 03:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____