
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000799
16:36 11/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:18
136.0 C 220kPa 4:18
136.3 C 224kPa 4:48
136.4 C 225kPa 5:18
136.2 C 223kPa 5:48
136.2 C 223kPa 6:18
136.4 C 224kPa 6:48
135.9 C 221kPa 7:18
136.3 C 223kPa 7:48
136.3 C 223kPa 7:48
VENTING 7:48
AIR DRYING 8:46
CYCLE COMPLETE 8:48

Digital Signature #
ABB1548709131681
