
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006253
14:39 06/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:46
STERILIZING 3:48
136.5 C 222kPa 3:48
136.0 C 222kPa 4:18
136.2 C 223kPa 4:48
136.4 C 225kPa 5:18
136.4 C 225kPa 5:48
136.1 C 220kPa 6:18
136.3 C 223kPa 6:48
136.4 C 225kPa 7:18
136.4 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:09
CYCLE COMPLETE 12:21

Digital Signature #
116CBA545750E6BB
