

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000812  
13:07 01/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 3:02  
STERILIZING 4:13  
136.0 C 228kPa 4:13  
136.3 C 231kPa 4:43  
136.4 C 230kPa 5:13  
136.4 C 233kPa 5:43  
136.4 C 232kPa 6:13  
136.2 C 232kPa 6:43  
136.2 C 232kPa 7:13  
136.3 C 232kPa 7:43  
136.3 C 232kPa 7:43  
VENTING 7:44  
AIR DRYING 8:33  
CYCLE COMPLETE 23:33

-----  
Digital Signature #  
50661871D65F6B0F  
-----