
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001286
16:27 06/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:43
136.0 C 221kPa 5:43
136.4 C 225kPa 6:13
136.3 C 224kPa 6:43
136.2 C 224kPa 7:13
136.3 C 224kPa 7:43
136.4 C 223kPa 8:13
136.3 C 224kPa 8:43
136.2 C 222kPa 9:13
136.2 C 222kPa 9:13
VENTING 9:13
AIR DRYING 10:11
CYCLE COMPLETE 10:46

Digital Signature #
74F5D1AEE8F141FE
