
Midmark M11 - v1.0.5
Total Cycles: 2802

Name: RITTER2
Sterilizer ID: V2313228
08/28/2023 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.8 F 0.0
2:00 178.7 F 0.0
4:00 175.3 F 0.6
6:00 196.0 F 4.7
8:00 215.5 F 8.8
10:00 225.4 F 11.3
12:00 232.4 F 9.7
14:00 235.8 F 9.4
16:00 244.3 F 13.3
18:00 243.6 F 12.3
20:00 256.1 F 18.8
22:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.7 F 29.5
2:00 271.6 F 28.9
3:00 272.4 F 29.1
4:00 271.9 F 29.2

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 209.2 F 0.0
10:00 215.5 F 0.0
15:00 212.5 F 0.0
20:00 209.9 F 0.0
25:00 209.9 F 0.0
30:00 208.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

08/28/2023 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____