
Midmark M11 - v1.0.0
Total Cycles: 2338

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 189.5 F 4.5
4:00 201.9 F 9.5
6:00 221.1 F 6.9
8:00 232.8 F 10.2
10:00 235.1 F 7.7
12:00 243.4 F 12.1
14:00 245.3 F 12.9
16:00 258.2 F 19.9
18:00 269.7 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.8 F 28.9
2:00 272.4 F 29.0
3:00 271.6 F 28.4
4:00 271.8 F 28.8

Min 271.3 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 235.3 F 0.0
10:00 235.0 F 0.0
15:00 229.3 F 0.0
20:00 226.3 F 0.0
25:00 227.8 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

09/19/2022 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____