

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3131

Name: RITTER 1  
Sterilizer ID: V2178198  
04/19/2023 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.0 F 0.0  
2:00 170.0 F 0.0  
4:00 201.4 F 3.6  
6:00 212.3 F 8.7  
8:00 225.3 F 9.9  
10:00 233.4 F 9.8  
12:00 236.8 F 8.7  
14:00 245.0 F 12.6  
16:00 245.6 F 13.0  
18:00 259.1 F 20.2  
20:00 270.6 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 28.9  
2:00 271.4 F 28.1  
3:00 272.7 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 223.6 F 0.0  
10:00 240.9 F 0.0  
15:00 228.7 F 0.0  
20:00 231.1 F 0.0  
25:00 229.4 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 20:46  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:55  
CYCLE COMPLETE

04/19/2023 02:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_